



Homeland Security

May 18, 2009

Beth Wann
Director, Mission Essentials Division
Contracting Officer
Office of Acquisition
Transportation Security Administration
701 S. 12th St.
Arlington, VA 22202

Beth,

The Trace Detection Laboratory at TSL has completed an evaluation of three consumable supply items from DSA Detection, as compared to performance of the OEM item, when tested under identical conditions. Item evaluated:

<u>DSA Part No.</u>	<u>Description</u>
CT1317 (25 ct.) CT1319 (100 ct.)	Calibration trap, multi-purpose, Itemiser 3/DX
ST1316 (25 ct.) ST1318 (100 ct.)	Sample trap, multi-purpose, Itemiser 3/DX
SIG7577	Inlet gasket, Smiths 400B

The test results reveal no significant differences between these DSA products and the equivalent OEM products.

A handwritten signature in cursive script that reads "Sheldon Brunk".

Sheldon Brunk, Ph.D.
Research Chemist, Trace Detection
Independent Test & Evaluation
Transportation Security Laboratory
Atlantic City Airport, NJ 08405